

Title:

Certificate of Analysis (CoA)

Date: 11/20/2025
Date Tested: 11/20/2025
Customer: Lifeways Research
Testing material: Semaglutide GLP-1
Lot Number: S20.001
BT Sample ID: 005000037140347
Labeled Peptide
Content/Potency: 20 mg
Storage: R.T.
Packing material: Clear glass vial with white label, silver crimp and red cap
Labeled as: Semaglutide GLP-1
Manufacturer: GLP-123
Testing Purpose: FTIR and HPLC analysis for the identification, purity, potency and composition of a peptide product. It does not provide information on particulate matter, microbial contamination or presence of endotoxins.

Test	Method	Specification	Result
General Appearance	USP <630>	White powder	White powder
Mass	USP <41>	As recorded	119.1 mg
FTIR Identification and Composition Analysis	USP <197A>	Sample spectrum should confirm the content of peptide via characteristic bands	FTIR sample spectrum confirms the presence of peptide with addition of excipient(s)/fillers such as mannitol.
HPLC Purity of Peptide Assay (UV)	USP <621>	Specifications: $\geq 98\%$	99.9%
HPLC Potency Assay (UV)	USP <621>	Specifications: 95 – 105% of 20 mg	24.2mg (120.8%)

The results of the CoA relate only to the item(s) tested and applied to the sample as received.

Kimberly Quinn, MS
Lab Manager
BTLabs


Andrea Castro, AS
Scientist-I
BTLabs

5730 Corporate Way | Suite 220 | West Palm Beach, FL 33407
Phone: (561) 625-0133
E-mail: info@btlabtesting.com | Website: <https://btlabtesting.com>


Title:

Certificate of Analysis (CoA)

FTIR ID and Composition Analysis: Semaglutide GLP-1 Lot S20.001



HPLC Purity and Potency Assay @ 280 nm: Semaglutide GLP-1 Lot S20.001



Semaglutide GLP-1 Lot S20.001 @ 280 nm

Peak #:	Retention Time (min)	Area (mAU*s)
1	9.402	1770.5